

DIGITAL CONDUCTIVITY METER



0- 1000 cm In 5 Range Digital Conductivity Meter

Specification :

- Conductivity Range: 0.1 μ S/cm to 1000 mS/cm
- Resolution: 0.01 μ S/cm
- Accuracy: ± 1 % F.S. ± 1 digit
- TDS range: 0 – 100 g/L
- Cell Constant: Acceptable from 0.1 to 5.0
- Temperature: 0 to 100 °C
- Temperature Resolution: 0.1 °C
- Temperature Compensation: Automatic (0-100 °C)
- Display: LCD with Backlight
- Accessories: Conductivity Cell, Temperature probe, Stand & Clamp
- Calibration: 1 point

NOTE: There may be any change in specification due to continuous R & D without notice.

VIJAYANTA TECHNOLOGIES PVT. LTD.

(Formerly Vijai Electronics)

Dr. Baldev Singh Marg 28/147 Civil Lines, Roorkee-247667 Distt. Haridwar, Uttarakhand

Phone No.: 01332 – 272509, 7579200827

E-Mail : vijayantatechologies@gmail.com, vijaielectronics1965@gmail.com