

LABORATORY CENTRIFUGE MACHINE



Laboratory Centrifuge Machine

Specifications:

- i) Max Speed : 6000 rpm
- ii) Max RCF :5070 g
- iii) Max. Capacity : 400 ml
- iv) Digital Timer Range :0-99 min
- v) Noise: \leq 60dB.

Rotor :-

Specifications:

- i) Max Speed : 6000 rpm
- ii) Max RCF :4700 g

(With spare angle Roter Head Complete)

NOTE: There may be any change in specification due to continuous R & D without notice.

VIJAYANTA TECHNOLOGIES PVT. LTD.

(Formerly Vijai Electronics)

Dr. Baldev Singh Marg 28/147 Civil Lines, Roorkee-247667 Distt. Haridwar, Uttarakhand

Phone No.: 01332 – 272509, 7579200827

E-Mail : vijayantatechologies@gmail.com, vijaielectronics1965@gmail.com